

## Giant – Comparative Arsenic Analytical Results

	Giant Assay Lab Colorimetric	DIAND Assay AA-Hydride
<b>Station 43-1      Final Decant</b>		
July 26	0.63	0.780
	0.65	0.765
		0.846
Sept 23	0.74	0.939
		0.954
		0.952
<b>Station 43-5      Baker Creek</b>		
July 26		0.711
		0.665
		0.620
Sept 23		0.961

Giant is reading As levels 15% to 28% low  
DIAND results confirmed by ICP

Colorimetric Procedure => Major Inteference from Antimony

## Giant – Comparative Nickel Analytical Results

	Giant Assay Lab Ni (AA)	DIAND Assay Ni (AA)
<b>Station 43-1      Final Decant</b>		
Sept 23	0.48	0.612
		0.624
		0.585

Copper analysis were also low but not to the same degree