

Giant – Comparative Arsenic Analytical Results

	Giant Assay Lab Colorimetric	DIAND Assay AA-Hydride
Station 43-1 Final Decant		
July 26	0.63	0.780
	0.65	0.765
		0.846
Sept 23	0.74	0.939
		0.954
		0.952
Station 43-5 Baker Creek		
July 26		0.711
		0.665
		0.620
Sept 23		0.961

Giant is reading As levels 15% to 28% low
DIAND results confirmed by ICP

Colorimetric Procedure => Major Inteference from Antimony

Giant – Comparative Nickel Analytical Results

	Giant Assay Lab Ni (AA)	DIAND Assay Ni (AA)
Station 43-1 Final Decant		
Sept 23	0.48	0.612
		0.624
		0.585

Copper analysis were also low but not to the same degree